



# LIGAMENT STAPLE



A	X	Y	Z
8mm	13mm	20mm	3.7mm
6mm	11mm	25mm	3.7mm
8mm	13mm	25mm	3.7mm
11mm	16mm	25mm	3.7mm
16mm	21mm	25mm	3.7mm

The Ligament staple is made up of titanium alloy. It can be used for attaching soft tissue to the bone. Low-profile, wide staple base provides better load distribution while impacting. Spike-post leg design encourages uninterrupted vascular flow to the underlying tissue. Sharp leg points for easier penetration into cortical bone. The ligament staples are available in different sizes. The staple impactor is provided in the instrument set for the staple holding and easier staple insertion into the bone.

## Ordering Information

Code	Product Description
464-8-13-20	Ligament Staple, 8mm x 13mm x 20mm, Titanium
464-6-11-25	Ligament Staple, 6mm x 11mm x 25mm, Titanium
464-8-13-25	Ligament Staple, 8mm x 13mm x 25mm, Titanium
464-11-16-25	Ligament Staple, 11mm x 16mm x 25mm, Titanium
464-16-21-25	Ligament Staple, 16mm x 21mm x 25mm, Titanium

## Recommended Instrumentation

Code	Product Description	Compatible Implants
7-014-76	Staple Impactor	464-8-13-20, 464-6-11-25, 464-8-13-25, 464-11-16-25, 464-16-21-25

STERILE EO