

AUXILOCK[®] AUXPIK SUTURE-BASED ANCHOR



AUXILOCK[®] AuxPik suture-based anchor is made with UHMWPE suture anchor body and PEEK OPTIMA eyelet tip. The anchor is designed to deliver efficiency and promote ease of use. The AuxPik suture-based anchor is available in 3.2mm diameter with various combination of BioBraid and suture tapes. It can be used for rotator cuff repair surgeries. The AUXILOCK[®] AuxPik suture-based anchor provides a small footprint and also asserts subcortical fixation for anchor insertion. The drill guide and anchor driver combination is well designed to improve the performance and reliability.

Ordering Information

Code	Product Description	Suture/Tape Colour
6-001-03	AUXILOCK [®] 3.2mm AuxPik Suture-Based Anchor With Two #2 BioBraid: White/Blue & White/Black	
6-001-09	AUXILOCK [®] 3.2mm AuxPik Suture-Based Anchor With One #2 BioBraid and One 1.4mm Suture Tape: White/Black & White/Blue	

Recommended Instrumentation

Code	Product Description	Compatible Implants
7-013-01	8 Point Slotted Drill Guide, Length 190mm	6-001-03, 6-001-09
7-013-13	Ø3.2mm X 200mm Drill Bit	6-001-03, 6-001-09
7-013-23	Ø3.2mm Awl	6-001-03, 6-001-09
7-013-02	8 Point Drill Guide, Length 100mm	6-001-03, 6-001-09

STERILE EO